MINISTRY OF AVIATION DLRD/RRE					
Specification MOA/CV6152	SECURITY				
Issue 1 Dated 1st April 1965	Specification	Valve			
To be read in conjunction with K1001	Unclassified	Unclassified			
TYPE OF VALVE - Cathode Ray Tube DEFLECTION - Magnetic FOCUS - Magnetic GUN - Triode, with beam BULB - Glass. External SCREEN - BE (Aluminium ba PROTOTYPE - VA9229	MARKING See K1001/4 BASE Modified B9A See drawing Page 8				
RATINGS (Absolute non-simultaneous and not f Heater Voltage Heater Current Max. Anode Voltage Max. Cathode Current Max. Heater/Cathode Voltage	(V) (A) (kV) (µA)	6.3 0.6 10 100	NOTES	CONN Pin 1 2 3 4 5 6 7 8	Electrode no connection grid g no connection heater h heater h grid g no connection cathode k
Heater negative Heater positive Min. Resolution (Square Wave) at 60% medulation (cycles	s/com)	100 75 150	A A	9 Side contact	cathode k grid g anode a
Typical Operating Condit Anode Voltage Cathode Current	(kV) (µA)	5 3			CONTACT ecessed contact ng page 8
Gg - all (max) Ck - all (max)	(pF) (pF)	6			ENSIONS ng on page 8
				La composito de la composito d	

NOTES

- A. Voltage may be temporarily increased to 250, either polarity on short term overload condition.
- B. NATO Stock No. 5960-99-037-4252

No deflection

CV6152

To be performed in addition to those applicable in K1001 Page 2 LIMITS K1001 Insp. Sym-Test Conditions Test Units Ref. Level bol Min. Max. (a) 5A.3.1 No Voltages. See General Inspection-100% Dimensions drawing on page 8 (b) 100% 5A.3.2.2 No Voltages Loose Particles (c) Capacitance 5% 5A.4.6 Cathode - all 6.0 1. pF 2. Grid - all 6.0 рF 3. Anode - External 400 $p\mathbf{F}$ coating ۷h Va (v) (V) (kV) (a) 5A.4.1.2 100% 10 Grid Insulation 2.0 Ig uA leakage Current Alternative Method 6.3 Adjust for 10 Increase in cut-off 100% ٧ 20 Cut-off voltage Resistor 25 megohms 5A.4.1.3 6.3 0 100% Heater-Cathode Ihk Heater at 200V Leakage Current (a) positive and 20 uA (b) negative with 20 uA respect to cathode (f) 6.3 0 Heater Current 100% Ιh 0.55 0.65 A For the following tests the beam centralising magnet is to be adjusted to align beam in the centre of the limiting aperture and no further adjustment is to be made. (g) Neg. Grid Voltage Vg1 100% 15 | 25 5A.4.3 6.3 Adjust for 5 ٧g V cut-off Record Vg1 Adjust for optimum focus

TESTS (Cont'd)

Page 3				IBBID (CONT C)					
K1001	Test Conditions			Manh.	Insp.	Sym-	LIMITS		Units
Ref.	те	est Conditio	ns	Test	Level	bol	Min.	Max.	Units
	Vh (V)	Vg (V)	Va (kV)	(h)					
5A.5.4	6.3	Adjust	5	Neg. Grid Voltage Vg2	100%		Reco	rd Vg2	v
	appli	ecting field led to give ssed raster	8	Grid Drive Vg1 - Vg2 Intensity to increase continuously from zero with increase of grid drive.	100%		3.0	5.5	v
	to pr metri 0.011 cande	Voltage adj roduce a pho ic intensity Orthochro ela Type B ote 6.	to-						
5A.6.4.2	No de	Any convenien value efflecting or sing fields		(j) Displacement of centre of spot from point of intersection of neck axis with screen	100%			4.0	mm
5A•5•7•3	No de focus Beam defin	Any convenien value effecting or sing fields. diameter vi ned by beam ting apertur	sibly	(k) Unfocused spot diameter	100%			9	mm
	appli	0 ecting field ied to give cused raster	2	(1) Total Cathode Current	100%	Ik	50	-	Aış

K1001	Test Conditions			m t	Insp.	Sym-	Limi	ts	Timed de a
Ref.	Te	ssc Conditions	3	Test	Level	bol	Min.	Max.	Units
5A.6.3 and 5A.3.5	Vh (V) 6.3	Vg (V) Any convenient value	Va (kV) 5	(m) Useful Screen Area. See drawing Page 8 Screen Blemishes Note 2	100%				
		ocused raster or whole scree		1. Area above 50 units- Number 2. Area 13 to 50 units- Separation 3. Area below 13 units- Ignore unless in sufficient number to cause perceptable darkening of the screen when viewed from a distance of 1 ft.			10	0	mm
				Glass Blemishes 1. Area above 50 units- Number 2. Area 13 to 50 units- Separation 3. Area below 13 units- Ignore unless in sufficient number to cause perceptable darkening of the screen when viewed from a distance of 1 foot.	100%		10	0	mm
5A.4.2.1 and 5A.4.2.3	6.3	spot just visible - Ik = approx	10	(n) Flashover and Stray Emission Note 3 Tube to be free from	100%				
	10 n No d	0.5 µA -heat cathode minutes. Reflecting or using fields	ı	sparking and field emission after the first 5 seconds, and for not less than 15 seconds, after tapping has ceased.					
	6.3	Adjust to value measured in (h)	5	(o) Spatial Frequency Response See Appendix A Measured as:-	100%				
	appl scar for	lecting fields lied to give a naed line adju- optimum focu- line direction	a usted s in		/cm x	100	60		%

TESTS (Cont'd)

K1001	Test Conditions			Test	Insp.	Sym-	Limits		Units
Ref.	1	est Condition	8	Test	Level bol		Min.	Max.	OHLUS
	appl scar for	Vg (V) Adjust to value measured in (h) ecting fields ied to give a ined line adju- optimum focus line direction	sted in	Screen Noise Number of parallel lines examined within the inspection area acceptable for noise. Note 5.	100%		10		
5A.7.2				External Pressure	QA				
5A.3.9					QA				
5A.3.7				(g), Spot Movement * Holding Period - 7 days Repeat Tests (1) and (g) (1)	100%		50 15	- 25	μA V

notes

- 1. Raster size 2.54 cm square.
 - Scan conditions: line frequency 8.53 kg/s
 frame frequency 33.3 c/s
 approximately 245 lines
- 2. The unit of area is one square one-thousandth of one inch.

 Two or more blemishes separated by a distance less than the major dimension of the largest shall be regarded as a blemish with dimensions equal to that of the group.
- 3. The tube shall be held with the screen horizontal and uppermost. It shall be viewed for 15 seconds in darkened conditions whilst the tube neek is tapped with an approved forked, rubber covered wooden mallet at a minimum rate of 4 taps per second.
- * See Amendment 2 for details

Notes continued over/

NOTES (Cont'd)

4. Scan conditions : frequency 10 c/s or less

Length of line 4 mm approximately in a direction parallel to the longer sides of the inspection area. After focussing the spot may be extended, or deflected at a frequency in excess of 100 kc/sec, in a direction

at right angles to the direction of scan.


Focus conditions: A focus coil of good quality with astigmatism

correction shall be used, positioned with its centre of focus 148 mm from the front ground land of the tube.

A Ferranti coil type FC5 is suitable.

The focus must be readjusted at each test position.

Area of screen to be inspected with tube mounted with the anode connection vertical and uppermost.



Readings to be taken $\overline{\text{in}}$ each of the areas marked A to F above, and the average of the six readings used for the purpose of the test.

5. Scan conditions: frequency 1 to 10 o/s


Length of line 46 mm in a direction parallel within 2 degrees to the major axis of the inspection area,

which shall be as in Note 4.

Focus conditions: Details of coil and position on tube to be as in Note 4.

Spot adjusted for optimum focus.

Screen noise is measured by viewing the screen with a photomultiplier, so positioned that no significant variation of response as the spot traverses the tube face is introduced by the viewing system. The output of the photomultiplier is amplified and displayed on an oscilloscope. After completion of test an indelible mark shall be made on the face of the cathode ray tube at either end of the central line parallel to the accepted test lines.



TYPICAL TRACE

A line is acceptable when the decrease in light output does not exceed 50% peak at any point along the trace.

Photomultiplier

General purpose type, S.11 photocathode about im diameter, with spectral response and characteristics similar to EMI type 9524B.

Oscilloscope - minimum requirements

The C.R.T. should have a long persistance phosphor.

Scan rate: 1 to 10 cycles/second
Scan duration: 0.08 second to 0.8 second.

Vertical deflection range of not less than 5 centimetres with sensitivity of 1 Volt/cm.

A suitable type is Telequipment S.51

6. The spectral response curve of a detector to measure radiant energy in terms of the unit "Orthochromatic Candela Type B" is given below.

The response corresponds to that of a Selenium Barrier Layer Photocell combined with a special filter. The Qualification Approval Authority will held a sample of this filter.

The calibration of the detector will compare to that defined in K1001 Section 5A.2.(12).


SPECTRAL DISTRIBUTION OF A PHOTOCRIL/COLOUR FILTER

COMBINATION FOR MEASURING ORTHOCHROMATIC CANDELLA

TYPE B 100 90 80 70 RELATIVE:SENSITIVITY 60 50 40 30 20 4600 4000 4200 4400 4800 5000


WAVELENGTH A

For note 7. see Amendment 2



DIMENSIONS ARE IN MILLIMETRES.

ANODE CONTACT TO LIE WITHIN THESE LIMITS .



APPENDIX A

Test equipment for the measurement of Spatial Frequency Response using square wave modulation

Referring to Fig. 1 the operation of the system is as follows. A trace displayed on the C.R.T. under test is imaged through a microscope objective on to a grating containing alternate opaque and transparent strips as shown in Fig. 2, the direction of scan being at right angles to the bars of the grating. The transmitted light is collected by a photomultiplier via a field lens which fecuses the aperture of the microscope objective on to the photomultiplier cathode. The output from the photomultiplier is fed directly to an oscilloscope as a 'Y' deflection, and the oscilloscope 'X' scan is locked to the scan applied to the C.R.T. under test. To reduce errors due to phosphor noise a high frequency spot wobble is applied in a direction parallel to the grating bars.

Mechanical Requirements

The objective, grating, field lens and photomultiplier are housed in a light tight container, and this should be mounted to allow movement relative to the C.R.T. face. To allow of fine adjustment of the magnification when calibrating, the distance between the microscope objective and the grating must be adjustable. The field lens is mounted immediately behind the grating, (Fig. 1), and the photomultiplier cathode is preferably positioned at approximately the same distance from the field lens as the microscope objective.

Microscope Objective

Nominal Magnification x5 - suitable lens with N.A. 0.12 is made by Beck, London. Catalogue Number 5229. Colour Code Violet.

Grating

The grating pattern is shown in Fig. 2 and contains one cycle at low frequency, and the remainder at a single high frequency. The grating is made by photographic reduction on to Ilford Formalith film from a master. An accurate master may be made by machining a sheet of black anodised aluminium, and filling the machined areas with matt white paint. To maintain the accuracy of the master on the final grating great care is needed in the photographic process. The width of the bars on the grating is a function of the test frequency and the magnification of the objective lens. On C.R.T. CV6152 the resolution is measured at 150 cycles/cm and the system magnification is X5, thus each bar width on the grating is 0.0066 ins.

Photomultiplier

General purpose type, S11 photocathode about 1 in. diameter, with spectral response and characteristics similar to EMI type 9524B.

Field Lens

A single element type 1.5 in. diameter, focal length 3 in. is satisfactory.

Oscilloscope - Minimum requirements

The C.R.T. should have a long persistance phosphor.

Scan rate: 1 to 10 cycles/second

Scan duration: 0.08 second to 0.8 second

CV 6152

APPENDIX A (Cont'd)


Vertical deflection range of not less than 5 centimetres with sensitivity of 1 volt/cm.

A suitable type is Telequipment S.51.

It has been found convenient to use the oscilloscope time base saw tooth waveform to drive a scan amplifier to deflect the spot on the C.R.T. on test.

Calibration of system


The photomultiplier is removed and a lamp placed in the plane of the photocathode. The projected aerial image of the grating is viewed through a microscope assembly fitted with a stage micrometer calibrated in 0.001 in. The distance of the microscope objective relative to the grating is adjusted until 30 bars, (15 black and 15 white), occupy 0.040 in. on the stage micrometer. The position of the objective is locked and the photomultiplier replaced.



APPENDIX A

FIG. 2

TYPICAL GRATING
(Not to scale)



ELECTRONIC VALVE SPECIFICATIONS

SPECIFICATION MOA/CV6152 ISSUE 1 DATED 1st APRIL, 1965.

AMENDMENT No. 1

Page 8 Outline Drawing

Cross out existing Outline Drawing and paste over new Outline Drawing dated 28th July, 1965 attached hereto.

September, 1965.

(319491)

T.V.C. for R.R.E.

APP (51/2

ELECTRONIC VALVE SPECIFICATIONS SPECIFICATION CV6152 ISSUE 1 DATED 1st APRIL, 1965 AMENDMENT No.2

Page 5

Add the following test which shall immediately precede the holding period.

K1001 Ref.	TEST CONDITIONS			TEST	INSP.	SYMBOL	LIMITS		UNITS	
KIOOT KGI		DI WHOIIION	L	1231	LEVEL STREET		MIN	MAX	ailis	
	Vh (V) 6.3	Vg (V) Adjust to give cathode current of 0-1 µA See Note 7	Va (kV) 5	(q) Spot Movement Duration	100%		- 20	0.05	mm hrs	

Page 7 Add Note 7 as follows:-

The focused spot shall be observed with a photomicrometer.

The field of view of the instrument shall be restricted that a spot movement of 0.05mm shall produce a decrease in the detector output of not less than 50%.

The output of the detector shall be fed to a suitable chart recorder having a response time not exceeding 250 m secs for full scale deflection.

The crt shall be lined up in the system. The tube EHT shall then be switched off for a period of not less than 15 minutes and then switched on again. The test commences at the instant EHT is restored and lasts for a period of not less than 20 hours.

It is important that all supplies are suitably stable such that there is no appreciable change in cathode current, focus, or spot position for the duration of the test.

A suitable photomicrometer can be constructed by modifying the instrument described in Appendix A. The grating is removed and is replaced by a limiting aperture of suitable diameter, and the output of the photomultiplier is fed to a chart recorder.

Ministry of Technology/RRE

1996 28 F7 68